



Third Committee Draft (3CD) – Marked version

Project: Revision of R 16-1:2002 (*see BIML note on p3*)
Title: R xxx *Non-invasive non-automated sphygmomanometers*
Part 3: Test report format

Date: 27 September 2019
Document number: TC18_SC1_P1_N020
Supersedes document: TC18_SC1_P1_N014

Project Group: OIML TC 18/SC 1/p 1
Convenership: P.R. China
Conveners: Ms. Can Wang

Circulated to P- and O-members and liaison international bodies and external organizations for:

Discussion at (date and place of meeting):

Comments by:

Vote (P-members only) and comments by: **27 December 2019**

Explanatory note

According to OIML B 6 *Directives for OIML technical work*, OIML publications shall be reviewed every five years after publication by the responsible TC/SC to decide whether they should be confirmed, revised, or withdrawn. The current (old) R 16, which TC 18/SC 1 is responsible for, was published in 2002, and it has been identified that there are a few technical conflicts between the new ISO/IEC standard and OIML R 16. To avoid different requirements worldwide on blood pressure instruments, the convener started the work on drafting R 16-1 *Non-invasive mechanical sphygmomanometers* after the project for revision was approved at the 43rd CIML Meeting held in October 2008 in Sydney.

During this work, the convener received dozens of comments from the Project Group's members and liaisons. Therefore, we wish to express our most sincere thanks for all experts' kindness. Many of these proposals have been accepted and published in this current version.

The main changes proposed to R 16-1 are the following:

- OIML R16-1 should be revised into three parts according to OIML B 6, and now OIML Rxxx-3 refers to *Part 3 – Test report format*;
- “*Mechanical sphygmomanometers*” has been replaced by “*Non-automated sphygmomanometers*” to clarify the main distinction between types of sphygmomanometers. This is also an argument consistent with the new ISO/IEC standards;
- Those terms that are no longer used have been deleted;
- “A manual system for applying and releasing pressure” has been replaced by “a pneumatic system” to harmonise the term and also to allow for electro-mechanical control;
- The metrological requirements on MPEs no longer distinguish between “the first time” and “in use”, and the value has finally been set at “0.4 kPa (3 mmHg)” in consideration of recent technical developments in health care;
- In consideration of environmental and health protection, the requirement on the internal diameter of the mercury tube has been deleted to encourage the reduction of the total mercury volume;
- Parts of safety requirements are clarified in agreement with, or according to, the new ISO/IEC standards;
- Verification no longer distinguish between “initial” and “subsequent”;

The current document is the third Committee Draft (3CD). It was drawn up on the basis of the comments received from PG members on the second Committee Draft circulated in December 2018.

Definitions and references related to the International Vocabulary of Metrology – Basic and general concepts and associated terms (VIM) have been modified according to the 2012 edition.

BIML note

The existing R 16 is published in two parts:

R 16-1 Non-invasive mechanical sphygmomanometers, and

R 16-2 Non-invasive automated sphygmomanometers.

These are being revised under two separate projects: TC 18/SC 1/p 1 and TC 18/SC 1/p 2 respectively.

The existing part numbering of R 16 is not consistent with current OIML practice. All Recommendations are now published with separate parts for the metrological and technical requirements (designated R xxx-1), test procedures (designated R xxx-2), and test report format (designated R xxx-3). To avoid confusion with the existing numbering of R 16, on completion of the two projects, the existing R 16-1 and R 16-2 will be replaced by two separate Recommendations with new numbers and each having three parts:

R xxx-1 Non-invasive non-automated sphygmomanometers – Metrological and technical requirements

R xxx-2 Non-invasive non-automated sphygmomanometers – Test procedures

R xxx-3 Non-invasive non-automated sphygmomanometers – Test report format

R yyy-1 Non-invasive automated sphygmomanometers – Metrological and technical requirements

R yyy-2 Non-invasive automated sphygmomanometers – Test procedures

R yyy-3 Non-invasive automated sphygmomanometers – Test report format

This CD has been re-numbered in line with this arrangement.

Contents

Foreword	5
1 Test review	9
2 Maximum permissible errors of the cuff pressure indication	10
3 Effect of storage on cuff pressure indication	11
4 Effect of temperature on cuff pressure indication	12
5 Air leakage of the pneumatic system	13
6 Pressure reduction rate for deflation valves	14
7 Rapid exhaust	14
8 Pressure indicating devices	14
9 Additional technical requirements for mercury manometers	15
10 Additional requirements for aneroid manometers	16
11 Mechanical safety	17
12 Abort a measurement	18
13 Tamper proofing	18
14 Tubing connectors	18
15 Electrical safety (This test is optional within the OIML Certificate System)	19
16 Durability of markings	19
Foreword	4
B.1 Test review (ordered according Rxxx 2)	7
B.2 Maximum permissible errors of the cuff pressure indication	8
B.3 Effect of storage on cuff pressure indication	9
B.4 Effect of temperature on cuff pressure indication	10
B.5 Air leakage rate of the pneumatic system	11
B.6 Pressure reduction rate for deflation valves	12
B.7 Rapid exhaust valve	12
B.8 Mechanical safety	12
B.8 Pressure indicating devices	13
B.9 Electrical safety (This test is optional within the OIML Certificate System)	14
B.9 Additional technical requirements for mercury manometers	14
B.10 Additional requirements for aneroid manometers	15
B.11 Mechanical safety	16
B.12 Abort a measurement	17
B.13 Tamper proofing	17
B.14 Tubing connectors	17
B.15 Electrical safety (This test is optional within the OIML Certificate System)	17

Foreword

The International Organization of Legal Metrology (OIML) is a worldwide, intergovernmental organization whose primary aim is to harmonize the regulations and metrological controls applied by the national metrological services, or related organizations, of its Member States. The main categories of OIML publications are:

International Recommendations (OIML R), which are model regulations that establish the metrological characteristics required of certain measuring instruments and which specify methods and equipment for checking their conformity; the OIML Member States shall implement these Recommendations to the greatest possible extent;

International Documents (OIML D), which are informative in nature and intended to improve the work of the metrological services;

International Basic Publications (OIML B), which define the operating rules of the various OIML structures and systems;

International Guides (OIML G), which are informative in nature and which are intended to give guidelines for the application of certain requirements to legal metrology.

OIML Draft Recommendations, Documents and Guides are developed by Technical Committees or Subcommittees which are formed by the Member States. Certain international and regional institutions also participate on a consultation basis. Cooperative agreements are established between the OIML and certain institutions, such as ISO and the IEC, with the objective of avoiding contradictory requirements; consequently, manufacturers and users of measuring instruments, test laboratories, etc. may simultaneously apply OIML publications and those of other institutions.

International Recommendations, Documents and Guides are published in French (F) and English (E) and are subject to periodic revision.

Additionally, the OIML participates in the publication of **Vocabularies (OIML V)** and periodically commissions legal metrology Experts to write **Expert Reports (OIML E)**. Expert Reports are intended to provide information and advice to metrological authorities, and are written solely from the viewpoint of their author, without the involvement of a Technical Committee or Subcommittee, nor that of the OIML. Thus they do not necessarily represent the views of the OIML.

This publication - reference OIML R 16-1, edition **201X (E)** - was developed by the OIML Technical Subcommittee TC 18/SC 1 Blood pressure instruments. It was approved for final publication by the International Committee of Legal Metrology in **201X** and supersedes OIML R 16-1:2002 (E).

OIML Publications may be downloaded from the OIML web site in the form of PDF files. Additional information on OIML Publications may be obtained from the Organization's headquarters:

Bureau International de Métrologie Légale

11, rue Turgot - 75009 Paris - France

Telephone: 33 (0)1 48 78 12 82

Fax: 33 (0)1 42 82 17 27

E-mail: biml@oiml.org

Non-invasive Non-automated Sphygmomanometers

Part 3: Test Report Format

Explanatory notes on the test report format

i General

This Test report format, which is informative with regard to the implementation of Rxxx-1 in national regulations, presents a standardized format for the results of the various tests and examinations to which a type of sphygmomanometer shall be submitted with a view to its approval as well as for the results of verification tests. The tests are listed in Rxxx-2.

It is recommended that all metrology services or laboratories evaluating types of sphygmomanometers according to OIML R 16-1 or to national or regional regulations based on OIML R 16-1 use this Test report format, directly or after translation into a language other than English or French.

It is also recommended that this Test report format in English or in French (or in both languages) be transmitted by the country performing these tests to the relevant authorities of another country, under bi- or multi-lateral cooperation agreements.

In the framework of the OIML Certificate System for Measuring Instruments, use of the Test report format is mandatory.

ii Page numbering and the use of report page formats

In addition to the sequential numbering at the bottom of each page, a space has been left at the top of each page for numbering the pages of reports established following this model. In particular, each test is reported individually on a separate page following the relevant format. For a given report, it is advisable to complete the sequential numbering of each page by indicating the total number of pages in the report.

Where required, pressure values in the Tables can be replaced by values expressed in kPa.

Where required, these forms can be copied and used several times in cases where the test in question has to be repeated under varying conditions.

iii Definitions and formula

For the purposes of this test report format, the following definitions and formula, taken from the International Vocabulary of Basic and General Terms in Metrology (VIM, 2012 edition) are used.

Non-invasive non-automated Sphygmomanometers

OIML Rxxx Edition 201X (E)

TEST REPORT

TYPE APPROVAL TEST REPORT

VERIFICATION TEST REPORT

(For verification purposes tick those fields which are appropriate for verification according to your national regulations or which are listed in 1.2 under the heading: Summary of test results for verification.)

Number of report:

Object:.....

Type:

Serial number:

Manufacturer's name and address:

.....

.....

Customer's name and address:.....

.....

.....

Date of receipt:

Date/period of measurement:

Date of report: Number of pages:

Issuing Institute's name and address:

.....

.....

Characteristic values (principle of measurement, measuring unit, measuring range, range of display):.....

.....

Additional devices (printer, interface etc.):

.....

Reference manometer (model, serial number, expanded uncertainty, calibration certificate):


.....

.....

Time measuring device (model, serial number, expanded uncertainty, calibration certificate):

.....

Stamp/signature:



1 Test review**1.1 Summary of test results for type approval**

Clause	Subject	Test result	OIML requirement	Passed	Failed
2	Maximum permissible errors of the cuff pressure indication				
3	Effect of storage on cuff pressure indication				
4	Effect of temperature on cuff pressure indication				
5	Air leakage of the pneumatic system				
6	Pressure reduction rate for deflation valves				
7	Rapid exhaust				
8	Pressure indicating devices				
8.1	Nominal range and measuring range				
8.2	Analogue indication - Scale				
8.3	Analogue indication - First scale mark				
8.4	Analogue indication - Scale interval				
8.5	Analogue indication - Scale spacing and thickness of scale marks				
8.6	Digital indication				
9	Additional technical requirements for mercury manometers				
9.1	Portable devices				
9.2	Device to prevent mercury from being spilled during use and transport				
9.3	Quality of the mercury				
9.4	Graduation of the mercury tube				
10	Additional requirements for aneroid manometers				
10.1	Scale mark at zero				
10.2	Zero				
10.3	Pointer				
10.4	Hysteresis error				
10.5	Durability of manometer				
11	Mechanical safety				
11.1	Resistance to shock for handheld sphygmomanometer				
11.2	Non-automated sphygmomanometers used during patient transport				

11.3	Non-automated sphygmomanometers containing a mercury manometer				
12	Abort a measurement				
13	Tamper proofing				
14	Electrical safety				
15	Tubing connectors				
16	Durability of markings				

1.2 Summary of test results for verification

Clause	Subject	Test result	OIML requirement	Passed	Failed
2	Maximum permissible errors of the cuff pressure indication				
5 –	Air leakage of the pneumatic system				

Note 1: The sequence of the different tests is arbitrary; it follows the sequence of the different clauses in the text. The sequence of testing is at the discretion of the person conducting the tests.

Note 2: To be considered as approved or verified, an instrument must have successfully passed all the applicable tests.

2 Maximum permissible errors of the cuff pressure indication

For the limits of temperature and humidity see **Rxxx-1** 5.1: the temperature should be between 15– °C and 25– °C, the relative humidity should be between 15–% and 85–%.

To find out the error of the cuff pressure indication proceed as follows (up and down runs) at three different temperatures: e.g. 15– °C and 15–% relative humidity, 20– °C and 60–% relative humidity and 25– °C and 85–% relative humidity.

Table 1 Example: Temperature 20– °C and...% relative humidity **Unit**
(mmHg)

pressure mmHg	1st reading		2nd reading		mean		deviation		hysteresis	
	up	down	up	down	up	down	up	down	1st reading	2nd reading
0	2	0	0	4	1	2	1	2	2	4
50	52	54	54	54	53	54	3	4	2	0
100	106	100	104	104	105	102	5	2	6	0
150										
200										
250										
column 1	column 2	column 3	column 4	column 5	column 6	column 7	column 8	column 9	column 10	column 11

maximum deviation: 5– mmHg _____ maximum hysteresis: 6– mmHg

Column 1 = values measured by the reference manometer

Column 2, 3, 4 and 5 = results of the measurement of the instrument under test

Column 6 = (column 2 + column 4) / 2

Column 7 = (column 3 + column 5) / 2

Column 8 = column 6 - column 1

Column 9 = column 7 - column 1

Column 10 = | column 2 - column 3 |

Column 11 = | column 4 - column 5 |

Table 2: Temperature ...°C and...% relative humidity — Unit

(mmHg)

pressure mmHg	1st reading		2nd reading		mean		deviation		hysteresis	
	Up	down	up	down	up	down	up	down	1st reading	2nd reading

Maximum deviation: Maximum hysteresis:

Note 1: The hysteresis error is the absolute value of difference between the indications of the instrument when the same pressure is reached by increasing or decreasing the pressure.

Note 2: The time between up and down run should not be less than 5 minutes at the maximum pressure. A time difference from the first run to the second run of one hour is recommended.

Is the maximum deviation of all the readings of the instrument under test and the reference manometer less than or equal to ±0.4 kPa (±3 mmHg)?

Yes → Passed

No → Failed

3 Effect of storage on cuff pressure indication

Determine the error after the storage for 24 h at a temperature of 20 °C and for 24 h at a temperature of 70 °C and a relative humidity of 85%.

Note: The measurements are to be performed before and after applying the test conditions, respectively:

- First measurement at 20 °C and 60% relative humidity before the test (refer to Table 2);
- Storage of the instrument under test for 24 hours at 20 °C and 85% relative humidity, immediately followed by storage of the instrument under test for 24 hours at 70 °C and 85% relative humidity;
- Second measurement at 20 °C and 60% relative humidity after the test.

Each measurement requires two readings. Calculate the deviation of the mean of the two readings after storage in Table 3 from the mean calculated in Table 2. The result should be within the error limits

mentioned below.

Table 3 –Measurement at 20_°C and 60_% relative humidity after storage at ~~±~~20_°C and 70_°C

Unit (mmHg_____)

pressure mmHg	1st reading after storage		2nd reading after storage		Mean after storage		deviation between the mean after storage and Table 2	
	up	down	up	down	up	down	up	down

Maximum deviation:

Is the maximum deviation of the cuff pressure indication (mean value), after storage at ~~±~~20_°C and 70_°C, less than or equal to ± 0.4 _kPa (± 3 _mmHg) compared to the mean values at 20_°C and 60_% relative humidity before storage?

Yes → Passed

No → Failed

4 Effect of temperature on cuff pressure indication

For each of the following combinations of temperature and humidity, condition the device for at least 3 h in the climatic chamber to allow the device to reach steady conditions.

Table 4 Temperature 10_°C and 85_% relative humidity

Unit

(mmHg_____)

pressure mmHg	1st reading		2nd reading		mean		deviation	
	up	down	up	down	up	down	up	down

--	--	--	--	--	--	--	--	--

Maximum deviation:

Table 5 Temperature 40_°C and 85_% relative humidity

Unit

(mmHg_)

pressure mmHg	1st reading		2nd reading		mean		deviation	
	up	down	up	down	up	down	up	down

Maximum deviation:

Is the maximum difference between the indicated pressure of the manometer of the device to be tested and the corresponding readings of the reference manometer at the relevant temperature value less than or equal to ± 0.4 kPa (± 3 mmHg), or ± 2 % of the reading, whichever is greater?

Yes → Passed
 No → Failed

5 Air leakage of the pneumatic system

Carry out the test over the whole measuring range at three equally spaced pressure steps at least (e.g. 7_kPa (50_mmHg), 20_kPa (150_mmHg), and 33_kPa (250_mmHg)). Test the air leakage rate over a period of 5_min (see Rxxx-2 4.2), and determine the measured value from this. Wait at least 60_s before reading each value.

Table 6

Unit

(mmHg_)

pressure	first reading	reading after 5 min	difference between the readings

--	--	--	--

Does the pressure drop over a period of 5 minutes correspond to an air leakage rate less than or equal to 0.5- kPa/min (4- mmHg/min)?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

6 Pressure reduction rate for deflation valves

Are the pneumatic system deflation valves capable of adjustment to a deflation rate of 0.3- kPa/s to 0.4- kPa/s (2- mmHg/s to 3- mmHg/s)?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

Are the pneumatic system deflation valves easily adjusted to these values.

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

7 Rapid exhaust

Time for the pressure reduction from 35- kPa to 2- kPa (260- mmHg to 15- mmHg):

$t_{Re} =$ s,

Is t_{Re} less than or equal to 10- s?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

8 Pressure indicating devices

8.1 Nominal range and measuring range

Does the nominal range extend from 0- kPa to at least 35- kPa (0- mmHg to at least 260- mmHg)?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

8.2 Analogue indication - Scale

Is the scale designed and arranged so that the measuring values can be read clearly and easily recognized?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

8.3 Analogue indication - First scale mark

Does graduation begin with the first scale mark at 0- kPa (0- mmHg)?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

8.4 Analogue indication - Scale interval

Is the scale interval 0.5_kPa or 2_mmHg for a scale graduated in kPa or mmHg, respectively?

Yes → Passed

No → Failed

Is each second scale mark indicated by greater length and is each fourth scale mark numbered in the case for a scale graduated in kPa?

Yes → Passed

No → Failed

Is each fifth scale mark indicated by greater length and is each tenth scale mark numbered in the case for a scale graduated in mmHg?

Yes → Passed

No → Failed

8.5 Analogue indication - Scale spacing and thickness of scale marks

Is the distance between adjacent scale marks not less than 1.0_mm and does the thickness of the scale marks not exceed 20_% of the smallest scale spacing?

Yes → Passed

No → Failed

Are the scale marks of equal thickness?

Yes → Passed

No → Failed

8.6 Digital indication

Is the digital scale interval 0.1_kPa (1_mmHg)?

If the measured value of a parameter is indicated on more than one display, do all the displays indicate the same numerical value?

Are numbers and characters clearly legible?

Yes → Passed

No → Failed

9 Additional technical requirements for mercury manometers**9.1 Portable devices**

Does the portable device have an adjusting or locking mechanism to secure it in a specified position of use?

Yes → Passed

No → Failed

9.2 Device to prevent mercury from being spilled during use and transport

Does the tube have a device to prevent the mercury from being spilled during transport and use and is this device efficient?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

Is the exhaust time taken for the mercury column to fall from 27_kPa to 5_kPa (from 200_mmHg to 40_mmHg) less than or equal to 1.5_s when the pressure in the system drops rapidly from 27_kPa to 0_kPa (from 200_mmHg to 0_mmHg)?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

9.3 Quality of the mercury

For type approval only:

Does the supplier confirm that the purity of the mercury is not less than 99.99_%?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

For type approval and for verification:

Does a visual inspection show a clean meniscus and no bubbles of air?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

9.4 Graduation of the mercury tube

Are the graduations permanently marked on the tube containing mercury?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

~~Note: Numbered at each fifth scale mark, the numbering shall be alternately on the right hand and left hand side of, and adjacent to, the tube.~~

10 Additional requirements for aneroid manometers

10.1 Scale mark at zero

If a tolerance zone is shown at zero, is it less than or equal to ± 0.4 _kPa (± 3 _mmHg)? Is it clearly marked? Is a scale mark at zero indicated?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

Note: Graduations within the tolerance zone are optional.

10.2 Zero

Is the movement of the elastic sensing element including the pointer unobstructed within 0.8_kPa (6_mmHg) below zero?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

Note: Neither the dial nor the pointer shall be adjustable by the user.

10.3 Pointer

Does the pointer cover between 1/3 and 2/3 of the length of the shortest scale mark of the scale?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

Is the pointer thinner than the scale mark at the place of indication and is the distance between the pointer and the dial less than or equal to 2_mm?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

10.4 Hysteresis error

Is the maximum hysteresis error throughout the pressure range less than or equal to 0.5_kPa (4_mmHg) according to Table 2?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

Note: The purpose of this test is to determine if the elastic sensing element has been exposed to a tension within the elastic range (i.e. the “Hooke’s” range) or not throughout the whole pressure range.

10.5 Durability of manometer

Is the change in the pressure indication after 10_000 alternating pressure cycles less than or equal to 0.4_kPa (3_mmHg) throughout the pressure range?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

11 Mechanical safety**11.1 Resistance to shock for handheld sphygmomanometer**

Does it have adequate mechanical strength when subjected to mechanical stress caused by normal use, pushing, impact, dropping, and rough handling?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

11.2 Non-automated sphygmomanometers used during patient transport

Does it have adequate mechanical strength when subjected to mechanical stress caused by normal use, pushing, impact, dropping, and rough handling?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

11.3 Non-automated sphygmomanometers containing a mercury manometer

Does it comply with the requirements of [Rxxx-1 5.1](#) after ~~testing-testing?~~

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

12 Abort a measurement

Is it possible to abort the blood pressure measurement at any time by activating the manual rapid exhaust valve, which shall be easily accessible?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

13 Tamper proofing

Tamper proofing of the manometer shall be achieved by requiring the use of a [special](#) tool.

Are means provided to prevent tampering and unauthorized access?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

Is it clear to an operator if tampering or authorized access has happened?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

14 Tubing connectors

~~Note: Users of equipment intended for use in environments employing intravascular fluid systems shall take all necessary precautions to avoid connecting the output of the blood pressure measuring device to such systems as air might inadvertently be pumped into a blood vessel if, for example, Luer locks were used. Luer lock connectors shall not be used with the tubing which connects the cuff to the manometer or measuring equipment, in order to avoid the possibility of inadvertent misconnection with other clinical systems.~~

Are Luer locks used?

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

~~Is the warning (see Note above and [Rxxx-1 7.5](#)) mentioned in the instruction manual?~~

Yes	<input type="checkbox"/>	→	Passed	<input type="checkbox"/>
<hr/>				
No	<input type="checkbox"/>	→	Failed	<input type="checkbox"/>

15 Electrical safety (This test is optional within the OIML Certificate System)

If the instrument is equipped with electrical devices, do these devices comply with regional and national safety regulations?

Yes → Passed

No → Failed

16 Durability of markings

If the markings fulfill the requirements?

The markings required shall be removable only with a tool or by appreciable force and shall be sufficiently durable to remain clearly legible during the expected service life of the sphygmomanometer.

Yes → Passed

No → Failed
